

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130302

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130302

SEPTEMBER 30, 2010

OPERATOR'S NAME:
MITCHELL BRIAN S
OPERATOR'S NUMBER: 4119
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME-INITIAL:
XXX
O.L. #: XXX
DEPT/AGENCY: KOTM
CASE/REPORT: XXX
TEST TYPE: V
ALCO TARGET VALUE: .002
ALCO 1/10: X124626

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.57 in
ALCO TARGET .001 13:08
BLANK TEST .000 13:09
INTERNAL STANDARD VERIFIED 13:09
ALCO TV 29.57 in .000 13:10
BLANK TEST .000 13:11
SUBJECT SAMPLE .000 13:11
BLANK TEST .000 13:12
ALCO TV 29.57 in .000 13:13
BLANK TEST .000 13:13

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130302

SEPTEMBER 30, 2010
TIME 13:14

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 42c
BAROMETER: 29.55 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*()+-./0123456789:;<=>?AB CDEFG
HIJKLMN O PQRSTU VWXYZ[\]^_`ab cde fghijklmno
pqrstuvwxyz{|}~"

VERIFICATION OF CALIBRATION REPORT

OCT 04 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130302</u> ✓	
Name <u>Brian Nuttall</u>		ID# <u>4119</u> ✓	Date <u>9-30-10</u> ✓
A	Agency <u>AST</u>	Phone # <u>486-4121</u>	
Instrument Location <u>Kodiak Post Interview Room</u>			
Alco S/N <u>x124626</u> ✓		Target Value <u>.082</u> ✓	High Pressure <u>250</u>
B	Alco Test Value Average <u>.088</u> ✓		<u>.088</u> ✓
	1 st Alco	2 nd Alco	
Signature <u>[Signature]</u>			
(OVER)			

BMB

10/5/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows (Date Noted within the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

